


## BID EVALUATION SHEET

Name of the Equipment: **SHORT TENDER OF ECG Machine**  
**KNOCK DOWN CRITERIA-(COMMERCIAL EVALUATION)** (To be evaluated by Purchase Department)  
 (All evaluation parameters defined below are mandatory for compliance)

SR. NO.	EVALUATION PARAMETERS	M/S Hospicare system Lahore	M/S Kasban International Lahore	M/S Radiant medical Pvt ltd Multan
1.	Original Receipt of Tender	Yes	Yes	Yes
2.	Technical Proposal	Yes	Yes	Yes
3.	Affidavit from Bidder according to bid form No.1	Yes	Yes	Yes
4.	CDR 2% estimated cost	Yes	Yes	Yes
5.	CNIC of authorized person	Yes	No	Yes
6.	Last year FBR Certificate	Yes	Yes	Yes
7.	Active NTN Number	0368358-3	2423601-2	2218935-1
8.	Past Performance	Yes	No	Yes
9.	Acceptance of terms and conditions certificate of bidding documents	Yes	Yes	Yes
10.	Distribution / authority letter	Yes	Yes	Yes
11.	Specification / detail information	Yes	No	Yes
12.	Country of manufacture USA, Europe, Japan (where applicable)	(Italy, Europe)	(Poland, Europe)	Japan
13.	Technical Engineering Capability	Yes	Yes	Yes
14.	Specification	<ul style="list-style-type: none"> <li>• Paper size maintained</li> <li>• Recording trace speed not</li> </ul>	<ul style="list-style-type: none"> <li>• Battery timing not confirmed from brochure</li> </ul>	<ul style="list-style-type: none"> <li>• 6 channel or 12 channel not mentioned</li> <li>• Variable gain not mentioned</li> </ul>

*[Handwritten Signature]*

REMARKS			
Not Responsive	<ul style="list-style-type: none"> <li>confirmed from brochure</li> <li>Variable gain not fall in given specs</li> <li>Model not available an website</li> <li>Battery timing not conformed</li> <li>Connectivity LAN optioned not standard</li> </ul>		
Not Responsive		<ul style="list-style-type: none"> <li>Paper size not mentioned</li> <li>Built in noise filter and baseline connection</li> <li>Capability to interface 2 x USB not LAN /WAN.</li> </ul>	
Not Responsive			<ul style="list-style-type: none"> <li>Paper roll not mentioned</li> <li>Built in Ac interference noise filter baseline connection not mentioned</li> </ul>



MISS. SARA FATIMA QAISRANI  
 BIO MEDICAL ENGINEER  
 TEACHING HOSPITAL, D.G. KHAN



DR. MUHAMMAD TAHIR ALI  
 PHARMACIST  
 TEACHING HOSPITAL, D.G. KHAN

PROF. DR. TEHSEEN IQBAL  
 PROFESSOR OF PHYSIOLOGY  
 D.G. KHAN MEDICAL COLLEGE D.G. KHAN  
 SENIOR TECHNICAL OFFICER