

**Bid Evaluation Sheet**  
**Package No.16(CARDIOLOGY)**

<b>Name of Equipment</b>	1	ECG Machine 6 Channel
	2	Defibrillator biphasic
	3	Pulse Oximeter with Paediatrics Sensor
<b>Participant bidders</b>	1-M/S Mediland Pakistan PVT LTD Lahore 2-M/S Medical Equipment & System Lahore 3-M/S Radiant Medical PVT LTD Lahore	

**Part -I & II**

**Knock Down Criteria (Commercial Evaluation )**

**All evaluation parameters defined below are mandatory for compliance**

Sr.No	Evaluation Parameters	M/S Mediland Pakistan PVT LTD Lahore	M/S Medical Equipment & System Lahore	M/S Radiant Medical PVT LTD Lahore
1	Complete Package / tender	YES	YES	YES
2	Original Receipt of Tender	YES	YES	YES
3	Affidivit from Bidder	YES	YES	YES
4	Bid Security	YES	YES	YES
5	Bid Validity	YES	YES	YES
6	Delivery Period	YES	YES	YES
7	Exclusive Authorization/ Sole Agent Certificate by Manufacturer	YES	NO	NO
8	Technical & Engineering Capability (As defined for the specific tender in specifications)/ List	YES	YES	NO
9	Certificate from manufacturer about the after sales services through agent or itself (in case specifically demanded in the sepcification)	YES	YES	YES
10	Vendor past performance (Incase of unsatisfactory performance detail must be mentioned)	YES	NO	NO
11	Availabiliy of Relevant Tools and Testing/ Calibration Enqupment	YES	YES	YES

12	Compliance of warranty as per tender	YES	YES	YES
13	International Standard	YES	YES	YES
14	Impted product on Current official website			

**PART – III**

**KNOCK DOWN CRITERIA - PRODUCT EVALUATION**

(All evaluation parameters defined below are mandatory for compliance.)

Sr.No	SPECIFICATION COMPLIANCE /EVALUATION PARAMETERS				
		<b>Reseller</b>	<b>M/S Mediland Pakistan PVT LTD Lahore</b>	<b>M/S Medical Equipment &amp; System Lahore</b>	<b>M/S Radiant Medical PVT LTD Lahore</b>
		<b>Brand</b>	<b>Philips Medical System</b>	<b>ZOLL</b>	<b>Bexen Cardio</b>
<b>1</b>	<b>Name of Equipment</b>	<b>Defibrillator biphasic</b>			
		<b>Model</b>	<b>DFM 100</b>	<b>M2</b>	<b>REMBEX 700</b>
<b>Country of Origin of Product/Model Number</b>			<b>Netherlan,China</b>	<b>USA,China</b>	<b>Spain</b>
<b>Compliance with defined quality standards</b>			<b>NO</b>	<b>Yes</b>	<b>NO</b>
<b>Specification Compliance features wise:</b>			<b>YES</b>	<b>NO</b>	<b>YES</b>
<b>Specifications:</b>			<b>YES</b>	<b>1-Specs Not confirmed from brochure and not fully routed</b>	<b>YES</b>
<b>Technical Eligibility of Product:</b>			<b>YES</b>	<b>Yes</b>	<b>YES</b>
<b>Technical Eligibility of Firm:</b>			<b>NO</b>	<b>NO</b>	<b>NO</b>
<b>BID STATUS:</b>			<b>Not Responsive</b>	<b>Not Responsive</b>	<b>Not Responsive</b>
		<b>Brand</b>	<b>NONIN Medical INC.</b>	<b>Medtronic Covidien AG group</b>	<b>MEDKINOSGUA</b>
<b>2</b>	<b>Name of Equipment</b>	<b>Pulse Oximeter with Paediatrics Sensor</b>			
		<b>Model</b>	<b>Avant 9700</b>	<b>PM 100N</b>	<b>SIGNSCOPE 5</b>
<b>Country of Origin of Product/Model Number</b>			<b>USA</b>	<b>USA,Korea</b>	<b>Lithunia</b>
<b>Compliance with defined quality standards</b>			<b>YES</b>	<b>NO</b>	<b>NO</b>
<b>Specification Compliance features wise:</b>			<b>YES</b>	<b>NO</b>	<b>YES</b>
<b>Specifications:</b>			<b>YES</b>	<b>YES</b>	<b>YES</b>
<b>Technical Eligibility of Product:</b>			<b>YES</b>	<b>NO</b>	<b>YES</b>
<b>Technical Eligibility of Firm:</b>			<b>YES</b>	<b>NO</b>	<b>NO</b>
<b>BID STATUS:</b>			<b>Responsive</b>	<b>Not Responsive</b>	<b>Not Responsive</b>
		<b>Brand</b>	<b>Philips Medical System</b>	<b>Cardioline</b>	<b>Welch Allyn</b>
<b>3</b>	<b>Name of Equipment</b>	<b>ECG Machine</b>			
		<b>Model</b>	<b>TC 10</b>	<b>100+</b>	<b>FU-230</b>
<b>Country of Origin of Product/Model Number</b>			<b>Netherlan,China</b>	<b>Italy</b>	<b>USA</b>
<b>Compliance with defined quality standards</b>			<b>YES</b>	<b>YES</b>	<b>NO</b>
<b>Specification Compliance features wise:</b>			<b>YES</b>	<b>NO</b>	<b>YES</b>

Specifications:	YES	1.Battery Timing not Confirmed 2.Z-Fold paper 3.Connecting Lead optional not standard	YES
Technical Eligibility of Product:	YES	NO	YES
Technical Eligibility of Firm:	NO	NO	NO
BID STATUS:	Not Responsive	Not Responsive	Not Responsive



