

Bid Evaluation Sheet
Package No.12 B: ANESTHESIA

Name of Equipment	1	Mobile Suction Machine Pump Heavy Duty
	2	Infusion Pumps
	3	Syringe Pump
Participant bidders	1-M/S Human Health Care Lahore 2-M/S Medical Equipment & System Lahore	

Part -I & II

Knock Down Criteria (Commercial Evaluation)

All evaluation parameters defined below are mandatory for compliance

Sr.No	Evaluation Parameters	M/S Human Health Care Lahore	M/S Medical Equipment & System Lahore
1	Complete Package / tender	YES	YES
2	Original Receipt of Tender	YES	YES
3	Affidivit from Bidder	YES	YES
4	Bid Security	YES	YES
5	Bid Validity	YES	YES
6	Delivery Period	YES	YES
7	Exclusive Authorization/ Sole Agent Certificate by Manufacturer	YES	YES
8	Technical & Engineering Capability (As defined for the specific tender in specifications)/ List	YES	YES
9	Certificate from manufacturer about the after sales services through agent or itself (In case specifically demanded in the specification)	YES	YES
10	Vendor past performance (Inc case of unsatisfactory performance detail must be mentioned)	YES	YES
11	Avallabiliy of Relevant Tools and Testing/ Calibration Enquipment	YES	YES
12	Compliance of warranty as per tender	YES	YES
13	International Standred	YES	YES
14	Improted product on Current official website		

[Handwritten signatures and initials]

KNOWLEDGE DOWN CRITERIA - PRODUCT EVALUATION

(All evaluation parameters defined below are mandatory for compliance.)

Sr.No SPECIFICATION COMPLIANCE /EVALUATION PARAMETERS

Sr.No	SPECIFICATION COMPLIANCE /EVALUATION PARAMETERS		M/S Human Health Care Lahore	M/S Medical Equipment & System Lahore
		Reseller		
		Brand	Medicinos GUS UAB	Fresenius
1	Name of Equipment	Infusion Pump		
		Model	MEDIFUSION IP	AGILIA VP MC
Country of Origin of Product/Model Number			Lithuania(Europe)	Europe China in letter of authorization
Compliance with defined quality standards			YES	YES
Specification Compliance features wise:			NO	YES
Specifications:			1-Door lock not confirmed from broucher	YES
Technical Eligibility of Product:			NO	YES
Technical Eligibility of Firm:			YES	YES
		Brand	Medicinos GUS UAB	Nouvag AG
2	Name of Equipment	Mobile Suction Machine Pump Heavy Duty		
		Model	PRIME 50	Vacuson 40
Country of Origin of Product/Model Number			Lithuania(Europe)	Switzerland
Compliance with defined quality standards			YES	YES
Specification Compliance features wise:			NO	YES
Specifications:			1-aspiration rate not compatible 2-Suction tubing of silicon with coupling connection for each jar(optional in broucher)	YES
Technical Eligibility of Product:			NO	YES
Technical Eligibility of Firm:			YES	YES
		Brand	Medicinos GUS UAB	Fresenius
3	Name of Equipment	Syringe Pump		
		Model	MEDIFUSION SP	AGILIA SP MC
Country of Origin of Product/Model Number			Lithuania(Europe)	Europe China in letter of authorization
Compliance with defined quality standards			YES	YES
Specification Compliance features wise:			NO	YES
Specifications:			1-drive accuracy +-2 instead of +-3	YES
Technical Eligibility of Product:			NO	YES
Technical Eligibility of Firm:			YES	YES
Demonstration			Not Produced/Not Satisfactory	Produced/Satisfactory
Package STATUS:			Not Responsive	Responsive