

Bid Evaluation Sheet
Package No.13: DENTAL

Name of Equipment	1	Dental Unit complete
	2	Dental X-Ray Machine
	3	X-Ray Development Equipment Automatic /OPG
Participant bidders	1-M/S Sigma International Lahore 2-M/S Combined Engineering Lahore 3-M/S Total Technologies PVT LTD Lahore	

Part-I & II

Knock Down Criteria (Commercial Evaluation)

All evaluation parameters defined below are mandatory for compliance

Sr.No	Evaluation Parameters	M/S Sigma International Lahore	M/S Combined Engineering Lahore	M/S Total Technologies PVT LTD Lahore
1	Complete Package / tender	YES	YES	YES
2	Original Receipt of Tender	YES	YES	YES
3	Affidavit from Bidder	YES	YES	YES
4	Bid Security	YES	YES	YES

21-04-2023

5	Bid Validity	YES	90Days instead of 180	YES
6	Delivery Period	YES	YES	YES
7	Exclusive Authorization/ Sole Agent Certificate by Manufacturer	NO	NO	Yes
8	Technical & Engineering Capability (As defined for the specific tender in specifications)/ List	YES	NO	YES
9	Certificate from manufacturer about the after sales services through agent or itself (in case specifically demanded in the specification)	NO	NO	YES
10	Vendor past performance (Incase of unsatisfactory performance detail must be mentioned)	NO	NO	Yes
11	Availability of Relevant Tools and Testing/ Calibration Enqupiment	YES	NO	YES
12	Compliance of warranty as per tender	YES	YES	YES
13	International Standred	NO	NO	YES
14	Improted product on Current official website			

PART – III

KNOCK DOWN CRITERIA - PRODUCT EVALUATION

(All evaluation parameters defined below are mandatory for compliance.)

Sr.No	SPECIFICATION COMPLIANCE /EVALUATION PARAMETERS
-------	---

		Reseller	M/S Total Technologies PVT LTD Lahore	M/S Combined Engineering Lahore	M/S Sigma International Lahore
		Brand	Dentsply Sirona	Cefla S.C Anthos (Italy)	Diplomat
1	Name of Equipment	Dental Unit complete			
		Model	Intego	Anthos A3 Plus (International)	DM20 + DC 180 + Xenos + D10L + KD 124 alongwith accessories
Country of Origin of Product/Model Number			Germany	Italy	Slovak Republic
Compliance with defined quality standards			YES	NO	NO
Specification Compliance features wise:			YES	NO	NO

✓

Specifications:


			YES		<p>1-Hand piece priority system not confirmed from brochure</p> <p>2-self generating LED Light not confirmed</p> <p>3-Anti-glare for the patient through clearly confined luminous spot not confirmed from brochure</p> <p>4-02x Working Stools model not confirmed</p>	<p>1-Hand piece priority system not confirmed from brochure</p> <p>2-self generating LED Light not confirmed</p> <p>3-Removable hand piece and tubing for cleaning and disinfection not confirmed from brochure</p> <p>4-Control of function keys of entry / exit position, tumblefilling, flushing the cuspidor and switch the light ON/OFF not confirmed from brochure</p> <p>5- Anti-glare for the patient through clearly confined luminous spot not confirmed from brochure</p> <p>6-02x Working Stools model not confirmed</p> <p>7- Electric Scalar (Built-in) not confirmed</p> <p>8-Slow Speed Air Motor not confirmed</p>
Technical Eligibility of Product:			YES	NO	NO	NO
Technical Eligibility of Firm:			YES	NO	Best X-DC	Orix HF Plus
	Brand		Xgenus			
2	Name of Equipment		Dental X-Ray Machine			
	Model		deGotzen	New Life Radiology		Ardet
Country of Origin of Product/Model Number			Italy	Italy		Italy
Compliance with defined quality standards			YES	NO		NO
Specification Compliance features wise:			YES	YES		YES


Specifications:

		<p>1-3D Upgradable with 3D Volume 8x8cm without stitching</p> <p>Not Upgradable to 3D as it comes in OPG / OPGwith CepH/ 3D OPG/ 3D OPG withCeph. Soonly Cepscan be upgraded.</p> <p>2-Automatic Patient positioning by Occlusal bite block and O2 laser beam system No automatic patient positioning with Occlusal bite block</p> <p>3- Image of individual quadrants not confirmed from brochure</p> <p>4-Image of upper or lower jaw not confirmed from brochure</p> <p>5- Artifact-reduced image not confirmed from brochure</p> <p>6-Thick-layer images of anterior region, both jaws not confirmed from brochure</p> <p>7-Multi slice in posterior tooth region not confirmed from brochure</p> <p>8- Multi-pulse High frequency generator: 100-120 KHz not confirmed from brochure</p>	<p>1-3D Upgradable with 3D Volume 8x8cm without stitching Not Upgradable to 3D as it comes in OPG / OPGwith CepH/ 3D OPG/ 3D OPG withCeph. Soonly Cepscan be upgraded.</p> <p>2-Automatic Patient positioning by Occlusal bite block and O2 laser beam system No automatic patient positioning with Occlusal bite block</p> <p>3- Image of individual quadrants not confirmed from brochure</p> <p>4-Image of upper or lower jaw not confirmed from brochure</p> <p>5- Artifact-reduced image not confirmed from brochure</p> <p>6-Thick-layer images of anterior region, both jaws not confirmed from brochure</p> <p>7-Multi slice in posterior tooth region not confirmed from brochure</p> <p>8- Multi-pulse High frequency generator: 100-120 KHz not confirmed from brochure</p>
Technical Eligibility of Product:	YES	NO	NO
Technical Eligibility of Firm:	YES	NO	NO
Package STATUS:	Responsive	Not Responsive	Not Responsive


M.A. J. R.

Dr.Farukh Hussain
Pharmacist Teaching Hospital
D.G.Khan
(Member)


Dr. Muhammad Asim Bhatti
Assistant Professor Surgery
Teaching Hospital D.G.Khan
(Member)


Miss. Sara Fatima Qasrani
Bio Medical Engineer
Teaching Hospital D.G.Khan
(Member)

Miss Rubab Iqbal
Director Finance
D.G Khan Medical College D.G Khan
(Member)


Dr. Muhammad Asif Gurmani
Assistant Professor Urology
D.G Khan Medical College D.G Khan
(Member)


Head of Department/End User
D.G.Khan Medical
College/Hospital D.G.Khan

Prof. Dr.Muhammad Azim Khan
Senior Technical Officer
Chairman