

Bid Evaluation Sheet **Package No.17(NEURO)**

Name of Equipment	1	Neuro physiological Monitor System
Participant bidders	1-M/S Radiant Medical PVT LTD Lahore 2-M/S Vertex Medical PVT LTD Lahore 3-M/S Medequips PVT LTD Lahore 4-M/S Shahco Medical PVT LTD Lahore	

Part-I & II **Knock Down Criteria (Commercial Evaluation)**

All evaluation parameters defined below are mandatory for compliance

Sr.No	Evaluation Parameters	M/S Radiant Medical PVT LTD Lahore	M/S Vertex Medical PVT LTD Lahore	M/S Medequips PVT LTD Lahore	M/S Shahco Medical PVT LTD Lahore
1	Complete Package / tender	YES	YES	YES	YES
2	Original Receipt of Tender	YES	YES	YES	YES
3	Affidavit from Bidder	YES	YES	YES	YES
4	Bid Security	YES	YES	YES	YES
5	Bid Validity	YES	YES	YES	YES
6	Delivery Period	YES	YES	YES	YES

[Handwritten signature]

7	Exclusive Authorization/ Sole Agent Certificate by Manufacturer	Same manufacturer and model is quoted by Shahco and Radiant	YES	YES	YES	Same manufacturer and model is quoted by Shahco and Radiant
8	Technical & Engineering Capability (As defined for the specific tender in specifications)/ List	YES	YES	YES	YES	YES
9	Certificate from manufacturer about the after sales services through agent or itself (in case specifically demanded in the specification)	YES	YES	YES	NO	YES
10	Vendor past performance (Incase of unsatisfactory performance detail must be mentioned)	No purchase order for quoted item attached	YES	YES	YES	NO
11	Availability of Relevant Tools and Testing/ Calibration Enquiment	YES	YES	YES	YES	YES
12	Compliance of warranty as per tender	YES	YES	YES	YES	YES
13	International Standard	YES	YES	YES	YES	YES
14	Improted product on Current official website	YES	YES	YES	YES	YES

PART – III

KNOCK DOWN CRITERIA - PRODUCT EVALUATION

(All evaluation parameters defined below are mandatory for compliance.)

SPECIFICATION COMPLIANCE / EVALUATION PARAMETERS						
Sr.No	Reseller	M/S Radiant Medical PVT LTD Lahore	M/S Vertex Medical PVT LTD Lahore	M/S Medequips PVT LTD Lahore	M/S Shahco Medical PVT LTD Lahore	
	Brand	Dr Langer	INOME	CADWEL	Dr Langer	
1	Name of Equipment	Neuro physiological Monitor System				
	Model	Avalanche Plus	ISIS XPRT	CASCADE IOMAX	Avalanche Plus	

Country of Origin of Product/Model Number	Germany CE	Germany CE	USA CE	Germany CE
Compliance with defined quality standards	Not Compliant	YES	Not fully comply	Not Compliant
Specification Compliance features wise:	1-only 32 channels available not upgrade of 64 channel 2-ECG option not available 3-required current is 250mA but available is only 100mA 4-required DNS/OCS channel is 2 but available is 1	YES	1-DNS not found 2-48 channels 3-upto 100mA while required 250mA	1-only 32 channels available not upgrade of 64 channel 2-ECG option not available 3-required current is 250mA but available is only 100mA 4-required DNS/OCS channel is 2 but available is 1
Specifications:				
Technical Eligibility of Product:	NO	YES	NO	NO
Technical Eligibility of Firm:	NO	YES	NO	NO
PACKAGE STATUS:	Not Eligible	Eligible	Not Eligible	Not Eligible


