

Bid Evaluation Sheet

Package No.2


Name of Equipment=Oxygen Generation System

Part - I & II

Knock Down Criteria (Commercial Evaluation)

All evaluation parameters defined below are mandatory for compliance

Sr.No	Evaluation Parameters	M/S ASK Associate Lahore	M/S Shahco PVT LTD	M/S G-Med PVT LTD Lahore	M/S Mediland Pakistan PVT LTD Lahore	M/S Radiant Medical Lahore	M/S KASBN International Lahore	M/Vertex Medical PVT LTD Lahore	M/S MATCO PVT LTD
1	Registered firm	YES	YES	YES	YES	YES	YES	YES	YES
2	Bid Security (2%)	YES	NO	YES	YES	NO	YES	YES	NO
3	One person One Bid	YES	YES	YES	YES	YES	YES	YES	YES
4	Manufacturer/Sole Distributor	NO (for major Components)	NO (for major Components)	NO (for major Components)	NO (for major Components)	YES	NO (for major Components)	NO (for major Components)	NO (for major Components)
5	Complete item with accessories	YES	YES	YES	YES	YES	YES	YES	YES
6	Certificate from manufacturer regarding Spare parts availability for 10 years	YES	YES	NO	NO	YES	YES	YES	YES
7	Comprehensive Warranty Period of 03-years	YES	YES	YES	YES	YES	YES	YES	YES
8	Bid Validity of required time	YES	YES	YES	YES	YES	YES	YES (06Months)	YES
9	Delivery Period within requirement	YES	YES	YES	YES	YES	YES	YES	YES
10	Remarks	Not Eligible	Not Eligible	Not Eligible	Not Eligible	Not Eligible	Not Eligible	Not Eligible	Not Eligible



SPECIFICATION COMPLIANCE / EVALUATION PARAMETERS										
Sr.No	Name of Equipment	Reseller	M/S ASK Associate Lahore	M/S Shahco PVT LTD	M/S G-Med PVT LTD Lahore	M/S Mediland Pakistan PVT LTD Lahore	M/S Radiant Medical Lahore	M/S KASBN International Lahore	M/Vertex Medical PVT LTD Lahore	M/S MATCO PVT LTD
1			OXYMAT	ERREDUE	SYSADVANCE	OXYWISE	PNEUMATECH (Atlas Copco)	SYSADVANCE	OMEGA AIR	BOGE
Oxygen Generation System										
Country of Origin of Manufacturer		SLOVAKIA	ITALY	PORTUGAL	SLOVAKIA	UK	Assorted	Assorted	Assorted	Assorted
Country of Origin of Product/Model Number		Assorted	Assorted	Not mentioned	Assorted	Assorted	Assorted	Assorted	Assorted	Assorted
Compliance with defined quality standards		not mentioned for major components	The wordings of tender copied as such. Not mentioned the exactly what standards does it meet. not mentioned for major components	The wordings of tender copied as such. Not mentioned the exactly what standards does it meet. not mentioned for major components	not mentioned for major components	Comply Substantially	not mentioned for major components	The wordings of tender copied as such. Not mentioned the exactly what standards does it meet. not mentioned for major components	not mentioned for major components	not mentioned for major components
Specifications/Sample Evaluation		Wrong quoted dryer; Refrigeration instead of desiccant dryer. The quoted system delivery is at Karachi major items of various manufacturer	Wrong quote for models of compressors (fixed speed type & desiccant dryer). Almost whole wordings quoted as a ditto of tender wordings which are contrary to their own models.	No models of the components mentioned and cannot be evaluated; vague offer Almost whole wordings quoted as a ditto of tender wordings which are contrary to their own models. Same model quoted by the M/S KASBN International	The system seems to be the combination of individual components which are difficult to integrate them to form a complete workable Oxygen generation plant system.	Comply Substantially	combination of individual components difficult to integrate (clarification is required in clear cut wordings that the complete installation shall be executed along with operations. (Mentioned that prices of such things are given in Fin. Offer) same model quoted by the M/S	Wrong quote for model of compressors (fixed speed type). The O2 tank quoted is of comp. air type. Almost whole wordings quoted as a ditto of tender wordings which are contrary to their own models.	Wrong quote for compressors (fixed speed type). Make of Cylinder Filling Station not given. Almost whole wordings quoted as a ditto of tender wordings which are contrary to their own models	Wrong quote for compressors (fixed speed type). Make of Cylinder Filling Station not given. Almost whole wordings quoted as a ditto of tender wordings which are contrary to their own models
Technical Eligibility of Product:		Not Eligible	Not Eligible	Not Eligible	Not Eligible	Eligible	Not Eligible	Not Eligible	Not Eligible	Not Eligible
Technical Eligibility of Firm:		Not Eligible	Not Eligible	Not Eligible	Not Eligible	Not Eligible	Not Eligible	Not Eligible	Not Eligible	Not Eligible
BID STATUS:		Non-Responsive	Non-Responsive	Non-Responsive	Non-Responsive	Non-Responsive	Not Responsive	Non-Responsive	Non-Responsive	Non-Responsive

SURGERY
NEUROSURGERY

- The system allows the patient's brain acquired from CT...
- The system allows to interactively visualize...
- The system allows to avoid critical structures...
- ...adjusting the virtual number of...

(Handwritten signatures and initials)